

التاريخ: 2026/09/20
الرقم: ص 2/2026/2015

رابطة الجامعيين / محافظة الخليل
جامعة بوليتكنك فلسطين
لجنة العطاءات المركزية
كراسة الشروط والمواصفات الفنية لعطاء
توريد اجهزة ومستهلكات لتخصص تكنولوجيا صناعة الاسنان
في كلية طب الاسنان
عطاء رقم: ص 2/2026/2015
ثمن الكراسة مبلغ (100 NIS)
(مائة شيكل)

استلام كراسة الشروط ابتداء من يوم الاحد الموافق 2026/09/20
تسليم الكراسة يوم الخميس الموافق 2026/10/01 الساعة الثانية عشرة ظهراً

مع أطيب أمنيات
مدير دائرة العطاءات والمشتريات المشتريات
ومقرر لجنة العطاءات والمشتريات المركزية
الاستاذ فراس موسى أبو شمسية
+970 599125751
direc-purch@ppu.edu



المحتويات :

١	١. اسم العطاء
٢	٢. الفهرس
٣	٣. الإعلان
٤	٤. ملاحظات عامة
٥	٥. تعهد والالتزام
٦	٦. تعليمات للمشاركين.
٨	٧. عموميات وجدول المواصفات والكميات والأسعار
١٧	٨. معلومات عن الشركة



إعلان عن عطاء
توريد أجهزة ومستهلكات لتخصص تكنولوجيا
صناعة الاسنان في كلية طب الاسنان
الرقم: 2015/2026/2



تعلن رابطة الجامعيين/ جامعة بوليتكنك فلسطين عن طرح عطاء توريد أجهزة ومستهلكات لتخصص تكنولوجيا صناعة الاسنان في كلية طب الاسنان ضمن الشروط والمواصفات الموضحة في كراسة وثائق العطاء، فعلى الشركات الراغبة بالدخول في العطاء إتباع الآتي: -

١. استلام الكراسة كاملةً من دائرة من العطاءات والمشتريات المركزية في مقر الجامعة/ ضاحية البلدية، مقابل دفع مبلغ (١٠٠ NIS) مائة شيكل غير مستردة تودع في حساب رابطة الجامعيين والجامعة رقم ٣٠٣٠٠ في البنك الإسلامي الفلسطيني اعتباراً من يوم الاحد الموافق ٢٠٢٦/٠٩/٢٠
٢. إرفاق شيك بنكي أو كفاله بنكية بقيمة ٥% من قيمة العطاء وبطرف منفصل، على أن تكون الكفالة البنكية سارية المفعول لمدة لا تقل عن تسعون يوماً.
٣. تسليم كراسة العطاء مع كافة التفاصيل بالطرف المختوم حتى الثانية عشرة ظهراً من يوم الخميس الموافق ٢٠٢٦/١٠/٠١ لدائرة العطاءات والمشتريات المركزية - رابطة الجامعيين.
٤. التوريد بأسرع وقت ممكن
٥. الأسعار بالشيك، شاملة ضريبة القيمة المضافة.
٦. الرجاء إرفاق صورة عن الفواتير الرسمية وشهادة خصم مصدر سارية المفعول مع العطاء.
٧. لجنة العطاءات غير ملزمة بقبول أقل الأسعار، وبدون إبداء الأسباب ويحق لها تجزئة العطاء. لمزيد من الاستفسار يمكن الاتصال مع الدكتور نور القواسمي ٠٥٩٥٤٤٤٤٧٥ خلال اوقات الدوام الرسمي.

ملاحظة: - أجور الإعلان على من يرسو عليه العطاء

مدير دائرة العطاءات والمشتريات
الأستاذ فراس موسى أبو شمسية

الملاحظات العامة لعطاء

توريد أجهزة ومستهلكات لتخصص تكنولوجيا صناعة الاسنان

في كلية طب الاسنان

الرقم: ص ٢/٢٠٢٦/٢٠١٥

يرجى مراعاة الآتي :-

١. يجب أن يكون المتقدم للعطاء شركة متخصص في مجال التوريدات الطبية.
٢. الأسعار بالشيك، شاملة ضريبة القيمة المضافة وخصم المصدر.
٣. الرجاء إرفاق شهادة خصم مصدر سارية المفعول مع العطاء
٤. الرجاء إرفاق الكاتالوجات ومدة الكفالة لكل جهاز من الأجهزة المطلوبة.
٥. تلتزم الشركة بالتوريد بأسرع وقت ممكن.
٦. الأسعار شاملة التوريد الى مباني الجامعة.
٧. لجنة العطاءات غير ملزمة بقبول أقل الأسعار، وبدون إبداء الأسباب.
٨. يكون السعر وفقاً للشروط الواردة في كراسة العطاء.
٩. الإعلان بالجريدة وكراسة الشروط الفنية للعطاء والاتفاقية وحدة واحدة وتقرآن معاً.
١٠. للجنة العطاءات الحق في زيادة أو إنقاص بعض الكميات
١١. للجنة العطاءات الحق في تجزئة العطاء
١٢. أجور الإعلان على من يرسو عليه العطاء.



تعهد وإقرار

أنا الموقع اسمي أدناه / قرأت الشروط واطلعت على المواصفات والبنود والتزمت بها التزاماً كاملاً وألتزم
بالأسعار المقدمة من قبلي لمدة تسعون يوماً من تاريخ فتح العطاء، وأتعهد بتقديم براءة ذمة "شهادة خصم
مصدر" من ضريبة الدخل سارية المفعول ومرفقة بالفاتورة الرسمية كما تعتبر هذه الثبوتيات أساساً لدفع
المستحقات اللازمة للمورد.

وبناءً على ذلك تمت المصادقة والتوقيع.

السادة / الشركة: _____

الشخص المسؤول: _____

رقم المشغل المرخص: _____

العنوان: _____

رقم الهاتف: _____

رقم الفاكس: _____

Emial : _____

التوقيع والخاتم

تعليمات للمشاركين بالعطاء

توريد أجهزة ومستهلكات لتخصص تكنولوجيا صناعة الاسنان

حضرات السادة : شركة _____ المحترمين

للمشاركة في العطاء ما يلي:-

1. تعتبر مقدمة كراسة الشروط والمواصفات وإعلان الجريدة جزءاً لا يتجزأ وتقران معاً.
2. يجب على صاحب العطاء التوقيع على وثائق العطاء كما يجب ختم العرض وكافة مرفقاته بخاتم صاحب العطاء.
3. لا يعتمد أي تعديل في الكراسة بسبب ما يدونه المتقدم من اشتراطات، ما لم تقبل بها لجنة العطاءات المركزية.
4. يجب على الشركة أن يضع أسعاره رقماً وكتابة على النموذج ويرفض أي عرض يحدث فيه المتقدم تشويشاً في أسعاره، واللجنة غير مسؤولة عن أية أخطاء قد يرتكبها المتقدم في وضع الأسعار.
5. مدة صلاحية الأسعار (٩٠) يوماً من ثاني يوم من فتح العطاء على الأقل.
6. على كل مناقص أن يرفق بالعطاء - لصالح رابطة الجامعيين - تأميناً للدخول في العطاء كفالة بنكية أو شيك مصدق من قبل البنوك المحلية بقيمة ٥% خمسة بالمائة من قيمة عرضه ولا ينظر في العروض الغير معززة بتلك التأمينات.
7. في حالة تأخير المورد عن الاعمال المحالة عليه تحسب غرامات التأخير بنسبة 1.5%، عن كل يوم تأخير، ومصادرة قيمة التأمين المرفق بالعطاء وقيده إيراداً للرابطة.
8. لاحقاً لبند رقم (٧) تقوم لجنة العطاءات المركزية بتنفيذ العطاء مباشرةً بالأسعار والشروط والطريقة المناسبة، من السوق المحلي مضافاً إليه (15%) من ذلك الفرق كنفقات إدارية.
9. عدم وجود أي تحفظات لها علاقة بسعر صرف العملات الأجنبية ويجب أن تكون الأسعار ثابتة حتى إتمام الالتزامات الفنية للمناقص في البند وصرف المستحقات وسيتم استبعاد أي عرض يوجد به تحفظات مرتبطة بأسعار صرف العملات الأجنبية.
10. يقدم العرض على النموذج أدناه ولا يحق إدخال أية تعديلات على وثائق العطاء. وإذا رغب الشركة تقديم ملاحظات أو عرض بديل عليه تقديم ذلك بمذكرة خاصة منفصلة شريطة تقديم العرض الأصلي كما هو، ولرابطة الجامعيين حق النظر بالمذكرة أو رفضها.
11. تكون المحاسبة وصرف جميع المستحقات للمناقص بعد الاستلام النهائي من لجنة الاستلام وعلى ضوء نتائج الفحص والاستلام حسب الأصول.

١٢. لا يجوز تحميل بند على بند آخر ولرابطة الجامعيين الخيار في إلغاء أي بند وتبقى أسعار البنود الأخرى ملزمة للمتزايد.
١٣. تعتبر الشروط العامة والفنية المطبقة في النظام العام للمشتريات جزءاً مكملًا لهذه الشروط في عطاءات رابطة الجامعيين.
١٤. يجوز للمناقص سحب عرضه بمذكرة موقعة منه وتودع في صندوق العطاءات قبل الموعد المحدد لفتح العطاء.
١٥. لا يجوز لصاحب العطاء التعديل أو المحو أو الطمس في قائمة الأسعار وأي تصحيح يجريه صاحب العرض عليها يجب إعادة كتابته رقماً وكتابة والتوقيع عليه وختمه.
١٦. إذا بلغت فئات الأسعار التي جرى عليها التعديل أو المحو أو الطمس أكثر من ١٠٪ من قائمة الأسعار جاز للجنة العطاءات والمشتريات المركزية استبعاد العرض.
١٧. يعتبر العرض المقدم من الشركة ملزماً له.
١٨. آخر موعد لتسليم العروض حتى الساعة الثانية عشرة ظهراً من يوم الخميس الموافق ٢٠٢٦/١٠/٠١.

❖ جدول الأسعار والمواصفات للأجهزة والمواد المستهلكة لتخصص تكنولوجيا
صناعة الاسنان:-

الرقم	الصنف المطلوب	الوحدة	الكمية	سعر الوحدة	الإجمالي
1	<u>Electric Casting machine:</u> ☐ Electrical induction or centrifugal/vacuum-pressure casting system suitable for dental laboratory applications. ☐ Suitable for casting Co-Cr and other common dental casting alloys. ☐ Melting temperature capability: minimum 1,500°C. ☐ Casting capacity: minimum 30–50 g per casting cycle. ☐ Compatible with standard dental casting crucibles and casting rings. ☐ Digital temperature and/or casting parameter control. ☐ Safety system with overheating and operating protection. ☐ Stable casting rotation/pressure system with reproducible casting results. ☐ Suitable for laboratory continuous operation and student training. ☐ Power supply: 220–240 V, 50 Hz or compatible with local electrical standards.	Machine	1		
2	<u>Porcelain furnace (ceramic firing furnace (Impress Ivoclar):</u> ☐ Combined ceramic firing and ceramic pressing furnace suitable for dental laboratory applications. ☐ Maximum firing temperature: at least 1,200°C. ☐ Suitable for pressing lithium disilicate and compatible pressable dental ceramics. ☐ Programmable firing cycles with adjustable: • Heating rate • Final temperature • Holding time • Vacuum level • Cooling parameters ☐ Integrated vacuum pump or compatible external vacuum system. ☐ Digital touch-screen or equivalent user interface.	Furnace	1		



				☐ Memory capacity for multiple firing and pressing programs. ☐ Automatic calibration and temperature control. ☐ Compatible with standard dental ceramic firing trays and pressing accessories. ☐ Suitable for continuous educational laboratory use. ☐ Power supply: 220–240 V, 50 Hz	
		1	Furnace	<u>Porcelain furnace (ceramic firing furnace(ivoclar):</u> ☐ High-quality dental ceramic firing furnace suitable for porcelain and ceramic restorations. ☐ Maximum temperature: at least 1,200°C. ☐ Programmable firing cycles. ☐ Adjustable heating rate and final temperature. ☐ Programmable vacuum firing function. ☐ Digital temperature control with high accuracy and stability. ☐ Memory for multiple ceramic firing programs. ☐ Suitable for firing dentine, enamel, opaque, stain, and glaze materials. ☐ Automatic calibration and temperature compensation. ☐ Suitable for standard dental ceramic firing trays. ☐ Power supply: 220–240 V, 50 Hz	3
		1	Printers	<u>3D printer:</u> ☐ Technology: LCD/MSLA or DLP resin 3D printing. ☐ Printing resolution: 35–50 µm XY resolution or better. ☐ Layer thickness: 10–100 µm or adjustable. ☐ Minimum build volume: approximately 190 × 120 × 200 mm or larger. ☐ Suitable for high-detail dental model printing. ☐ Compatible with validated dental model and laboratory resins. ☐ Touch-screen or user-friendly control interface. ☐ USB/Wi-Fi/LAN connectivity. ☐ Automatic or assisted platform leveling. ☐ Resin tank suitable for repeated laboratory use.	4



				☐ Open or widely compatible resin system preferred. ☐ Complete with build platform and required accessories. ☐ Power supply: 220–240 V, 50 Hz.	
		1	Machine	<u>Intraoral Scan:</u> ☐ High-accuracy intraoral optical 3D scanning system. ☐ Full-arch scanning capability for maxillary and mandibular arches. ☐ Accurate digital bite/occlusion registration. ☐ Scan accuracy suitable for restorative and laboratory CAD/CAM applications. ☐ Real-time 3D visualization. ☐ Color scanning preferred. ☐ Scan export formats: STL, PLY, and/or OBJ. ☐ Open data export preferred without mandatory proprietary file format. ☐ Compatible with dental CAD/CAM software, including exocad or equivalent. ☐ Anti-fogging scanner tip or automatic fog control. ☐ Autoclavable/reusable scanner tips or manufacturer-approved sterilization system. ☐ USB or equivalent high-speed computer connection. ☐ Suitable for continuous educational laboratory and clinical demonstration use. ☐ Complete with computer interface/software and required accessories.	5
		2	Machine	<u>Vacuum former:</u> ☐ Professional dental vacuum thermoforming machine. ☐ Suitable for thermoplastic sheets of approximately 0.5–5.0 mm thickness. ☐ Compatible with standard dental thermoforming sheets. ☐ Heating temperature control suitable for various thermoplastic materials. ☐ Vacuum pressure sufficient for accurate adaptation of sheets to dental models. ☐ Digital or adjustable heating and vacuum controls. ☐ Suitable for sheets approximately 125 mm diameter or equivalent standard dental size.	6



				☐ Fast and uniform heating system. ☐ Integrated or external vacuum system. ☐ Suitable for repeated student laboratory use. ☐ Power supply: 220–240 V, 50 Hz	
		7	VERSION	<u>CAD SOFTWARE(EXOCAD):</u> ☐ Professional dental CAD software suitable for laboratory applications. ☐ Latest stable version available at the time of supply. ☐ Compatible with Windows 11 Pro 64-bit. ☐ Compatible with common intraoral and laboratory scanner file formats. ☐ Support for STL, PLY, OBJ or equivalent open formats. ☐ Crown and bridge design module. ☐ Implant and abutment design capabilities. ☐ Removable partial denture design module preferred. ☐ Complete denture design capabilities preferred. ☐ Digital wax-up and diagnostic design. ☐ Virtual articulator and occlusion analysis. ☐ Automatic tooth arrangement and library functions. ☐ Export capability for milling and 3D-printing workflows. ☐ Compatible with commonly used dental milling machines and 3D printers. ☐ Educational/laboratory licenses sufficient for student training. ☐ Includes software updates and technical support according to the offered license	7
		2	Tool	<u>Surveyor:</u> ☐ Professional mechanical dental laboratory surveyor. ☐ Stable and rigid metal construction. ☐ Adjustable surveying arm with smooth vertical movement. ☐ Adjustable model table capable of tilting and rotating. ☐ Secure model-holding mechanism. ☐ Suitable for standard dental casts and removable partial denture models. ☐ Complete with standard surveying accessories, including: Analyzing rod	8



				- Carbon marker - Undercut gauges - Wax trimmer/parallel guide - Mandrel or equivalent accessories ☐ Fine adjustment mechanism for accurate surveying. ☐ Suitable for repeated undergraduate laboratory use.	
		2	Tool	<u>Minimajor:</u> Professional dental laboratory pressure curing unit suitable for cold-cure acrylic polymerization.	9
		10	علبة	<u>Cold cure acrylic (polymer +monomer):</u> Self-curing / cold-cure dental acrylic resin system consisting of polymer powder and monomer liquid, suitable for fabrication and repair of removable dental prostheses and laboratory training application	10
		40	علبة	<u>Light cure acrylic (mega tray):</u> Light-curing acrylic resin sheet/material for fabrication of individual/custom impression trays, suitable for dental laboratory use and compatible with standard light-curing units.	11
		250	KET	<u>Acrylic teeth (acry rock):</u> Prefabricated acrylic denture teeth, suitable for complete and partial removable dentures, with anatomically shaped anterior and posterior teeth and adequate wear resistance	12
		50	علبة	<u>Modeling wax(AARC DENTAL):</u> High-quality dental modeling/baseplate wax for fabrication of denture bases, wax rims, patterns, and laboratory wax-up procedures. Easy to adapt, carve, and manipulate without excessive brittleness.	13
		4	STICK	<u>POLISHING COMPOUND FOR DENTURE(ROUGE):</u> Dental laboratory polishing compound (rouge) suitable for final polishing of acrylic denture bases and other acrylic prosthetic appliances, producing a smooth and high-gloss surface	14



		20	CUP	<u>DISPOSABLE CUPS:</u> Disposable laboratory mixing cups suitable for mixing dental materials such as acrylic resin, gypsum, alginate, and other laboratory material	15
		40	لفة	<u>SS WIRE (0.6/0.7):</u> Stainless steel orthodontic/dental laboratory wire, round cross-section, available in 0.6 mm and 0.7 mm diameters, suitable for removable orthodontic appliances and dental laboratory application	16
		40	لفة	<u>SS WIRE (0.5/0.8):</u> Stainless steel orthodontic/dental laboratory wire, round cross-section, available in 0.5 mm and 0.8 mm diameters, suitable for fabrication of clasps, springs, retainers, and other removable appliance	17
		20	MOLD	<u>SILICONE MOLD FOR PARTIAL DENTURE (CLASS 1,2,3):</u> Reusable silicone training molds representing Kennedy Class I, Class II, and Class III partially edentulous arches, suitable for teaching and practical training in removable partial denture fabrication	18
		50	SHEET	<u>THERMOPLASTIC SHEET (RETAINER AND NIGHT GURD):</u> Thermoplastic dental sheets suitable for fabrication of orthodontic retainers and occlusal/night guards using a vacuum or pressure thermoforming unit. Sheets shall be dental-grade and suitable for intraoral use	19
		4	علب	<u>COBELT CROMIUM ALLOY:</u> Dental cobalt-chromium alloy suitable for casting removable partial denture frameworks and other dental laboratory applications. Alloy shall be compatible with conventional dental casting procedures and suitable for laboratory use	20
		4	كرتونة	<u>PHOSPHATE BONDED INVESMENT:</u> Phosphate-bonded dental casting investment material suitable for high-temperature casting of cobalt-chromium	21



				removable partial denture frameworks and dental alloy	
		3	لفة	<u>SPRUE FORMER (1.5,2 mm):</u> Dental laboratory sprue wax/wire formers with approximately 2.0–1.5 mm diameter, suitable for attaching wax patterns to casting sprue systems for dental casting procedures.	22
		10	علبة	<u>DENTAL PORCELAIN (OPAQUE)(CERAMCO):</u> Dental ceramic opaque porcelain powder compatible with the Ceramco ceramic system, suitable for masking metal substructures and establishing the basic shade before application of dentine porcelain.	23
		10	علبة	<u>DENTAL PORCELAIN (DENTINE) (CERAMCO):</u> Dental ceramic dentine porcelain powder compatible with the Ceramco ceramic system, suitable for building the main body and dentinal anatomy of porcelain-fused-to-metal restorations.	24
		4	علبه	<u>GLAZE (POWDER AND LIQUID):</u> Dental ceramic glaze system consisting of glaze powder and compatible glaze liquid, suitable for final glazing of dental porcelain restorations to achieve a smooth, natural-looking and glossy surface.	25
		10	علبه	<u>Dental Porcelain (Enamel):</u> Dental enamel/incisal porcelain powder suitable for reproducing the translucent enamel/incisal portion of ceramic restorations and for layering over dentine porcelain.	26
		5	علبه	<u>Opaque Base-Paste:</u> Dental ceramic opaque base paste suitable for application to metal substructures before opaque porcelain layering, providing uniform coverage and improved bonding/masking of the underlying metal.	27
		5	علبه	<u>Opaque Liquid:</u> Liquid medium specifically designed for mixing and application of dental ceramic opaque material, providing suitable consistency for controlled application to	28



				dental frameworks.	
		10	علبه	<u>Soprano 10 Dentine:</u> Dentine ceramic porcelain powder compatible with the Soprano 10 ceramic system, suitable for dentine layering and anatomical build-up of dental ceramic restorations.	29
		10	علبه	<u>Soprano 10 Enamel:</u> Enamel/incisal ceramic porcelain powder compatible with the Soprano 10 ceramic system, suitable for reproducing enamel translucency and incisal effects in ceramic restorations.	30
		2	علبه	<u>Soprano 10 Glaze:</u> Glaze material compatible with the Soprano 10 ceramic system, suitable for final surface glazing of ceramic restorations to provide smoothness, surface luster, and natural appearance.	31
		2	علبه	<u>Porcelain Liquid (Modeling Liquid):</u> Dental porcelain modeling liquid suitable for mixing ceramic powders and controlling consistency during porcelain build-up and layering procedures.	32
		20	block	<u>ZIRCONIA BLOCKS (AMANN GIRRBACH):</u> Standard milling block dimensions compatible with the Amann Girrbach milling system used by the dental laboratory, including SIZE (14MM/16MM/18MM) (SHADE A1,A2,B1)	33
		10	Blocks	<u>PMMA blocks:</u> Thickness Options: Approximately 14 mm, 16 mm, 18 mm, and 20 mm, depending on the required restoration and milling system (SHADE A1,A2,B1)	34
		2	kit	<u>ZIRCONIA STAINS kit:</u> Dental zirconia characterization stain kit suitable for internal/external staining and color characterization of zirconia restorations before final sintering and/or glazing. Kit shall include compatible stain materials and required mixing liquid(s).	35



		10	ingot	<u>LITHIUM DISILICATE:</u> Ingots shall be compatible with the dental pressing furnace/system available in the laboratory and supplied with the manufacturer's recommended pressing parameters.	36
		4	علبه	<u>Temporary resin (tempovac):</u> temporary/provisional resin material suitable for fabrication of temporary crowns and bridges using compatible FOR 3D PRINTER(1L)	37
		10	علب	<u>Model resin (tempovac):</u> Dental model resin suitable for fabrication of accurate dental working models using compatible FOR 3D PRINTER(1L)	38
		4	علب	<u>Castable resin (matrix):</u> Material shall provide clean burnout with minimal residue and be suitable for casting dental alloy(1L)	39
		20	Mold	<u>Preformed Silicon mold fixed prosthodontics training (bridge preparation):</u> Reusable preformed silicone training mold representing prepared teeth for fixed prosthodontic bridge fabrication and preparation exercises. Mold shall be suitable for repeated student laboratory training and practice.	40
		10	علبة	<u>ALCOHOL (70%):</u> 70% ethyl alcohol or isopropyl alcohol suitable for dental laboratory surface disinfection, cleaning, and general laboratory use. Supplied in sealed containers with appropriate labeling.	41
		10	saw	<u>Model saw (large size):</u> Large-size dental laboratory model saw with durable cutting blade, suitable for sectioning and trimming dental stone casts and working models with controlled and accurate cuts.	42
		10	علبة	<u>HEAT CURE monomer:</u> Dental acrylic monomer liquid (methyl methacrylate-based) suitable for mixing with compatible acrylic polymer powder for fabrication, , and laboratory processing of acrylic dental prostheses.	43
		5	kilo	<u>clear acrylic powder:</u> Heat-cure/self-cure clear acrylic polymer	44



				powder suitable for fabrication of transparent or clear dental laboratory appliances, bases, and other acrylic components. Must be compatible with the corresponding acrylic monomer.	
		50		expansion screw : Orthodontic expansion screw designed for incorporation into removable orthodontic appliances, suitable for controlled transverse expansion of dental arches during laboratory fabrication and orthodontic training.	45
		20	tool	Preformed Silicone Training Mold – Prepared Central, Canine, and Molar Teeth: Reusable, high-quality preformed silicone dental training mold containing anatomically shaped prepared central incisor, canine, and molar teeth . The central incisor, canine, and molar shall be pre-prepared according to standard fixed prosthodontic tooth preparation principles and suitable for crown and bridge training.	46
		20	Tool	RPD Mold	47
	المجموع الإجمالي بالشيكال				**

..... المبلغ بالحروف

خاص بالموارد - :

..... اسم الشركة:

..... رقم هاتف : رقم فاكس : جوال رقم :

..... التوقيع :

ختم الشركة الرسمي